

Channel	Part No.	Volume Range (µl)	Volume Increments (µl)	Test Volume (µl)	Accuracy (±%)	Precision (≤%)	Tip Spacing (mm)
4	4743	10 – 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 – 33.0
	4744	50 – 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 – 33.0
6	4763	10 – 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 – 19.8
	4764	50 – 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 – 19.8
8	4721	0.5 – 12.5	0.01	1.25	10.00	6.00	4.5 – 14.0
				6.25	4.00	1.60	
				12.5	2.00	0.80	
	4726	2 – 50	0.05	5	5.00	2.50	4.5 – 14.0
				25	2.50	0.60	
50	1.50	0.40					
4722	5 – 125	0.1	12.5	3.75	1.50	4.5 – 14.0	
			62.5	2.50	0.70		
125	1.60	0.35					
4723	10 – 300	0.5	30	4.00	1.20	9.0 – 14.0	
			150	2.00	0.60		
300	1.60	0.35					
4724	50 – 1250	1	125	6.00	1.10	9.0 – 14.0	
			625	2.40	0.50		
1250	1.20	0.30					
12	4731	0.5 – 12.5	0.01	1.25	10.00	6.00	4.5 – 9.0
				6.25	4.00	1.60	
				12.5	2.00	0.80	
4736	2 – 50	0.05	5	5.00	2.50	4.5 – 9.0	
			25	2.50	0.60		
50	1.50	0.40					
4732	5 – 125	0.1	12.5	3.75	1.50	4.5 – 9.0	
			62.5	2.50	0.70		
125	1.60	0.35					

1. The multichannel specifications represent the performance of all channels of a pipette.

2. Precision = coefficient of variation